

_FILE NO: DQA/DEL/MEC/GeM/TP/130

Sub Product Category -Injection and aspiration needles and accessories UNSPSC CODE : 421425.

Product category - Anaesthesia needles(Insulated needles for plexus anaesthesia and nerve blocks) UNSPSC CODE : 42142502. ITC-HS Code: 90189099

(* Demand aggregation to be done on UNSPSC code)

Sno	Parameters	Type of Field	Value 1	Value2	Value3	Value4	Remarks	Whether Filter required	Priority	unit	Validation
1	Application of needle	CH	Plexus Anaesthesia	Plexus nerve block				YES	1		
2	Needle suitable for ultrasound guided application	CH	YES	NO				YES	2		
3	Size of needle(Diameter)	N	20G	21G	22G	24G		YES	3	Gauge	
4	Length of needle	N	40	50	70	90		YES	4	mm	
			100	120	150						
5	Needle shall have precise thin transparent polymer insulation coating for better insulation of the needle and easy insertion.	CH	YES								
6	Inner lumen shall be insulated for better flow of local anaesthesia drug.	CH	YES								
7	Needle shall have clearly visible penetration marking in centimeter on the needle.	CH	YES								
8	Needle shall have reflectors at the tip of needle for precise visibility under ultrasound.	CH	YES	NO							
9	Sterilizing	CH	YES								
10	Packing	CH	Sterile, individually packed in a soft wrap .								
11	Certification available	CH					Indicate the availability of certificate like ISO/European-CE/US-FDA				Must declare

Approved By

Moderated By

Moderated By

Prepared By

(Sh.P.Singhal)

(Sh.S.K.Agrawal)

(Sh.A.K.M.Kashyap)

(S.K.NARESH)

DDG(Q.A.)(H.QRS)

D(Q.A.)

DD(Q.A.)

AD(Q.A.)-II