

FILE No: DQA/DEL/MEC/GeM/TP/102 (Minivac Suction Drain)

Major Product Category -WOUND CARE PRODUCTS, UNSPSC Code : 4231,

Sub Product Category -NEGATIVE PRESSURE WOUND CARE THERAPY PRODUCTS UNSPSC CODE : 423126

Product category - NEGATIVE PRESSURE WOUND CARE THERAPY SYSTEM OR KITS(MINIVAC SUCTION DRAIN)UNSPSC CODE : 42312602

ITC-HS Code : 90189099.

(*Demand aggregation to be done on UNSPSC code)

S No.	Parameters		Value 1	Value 2	Value 3	Value 4	Value 5	Value 6	Value 7	Remarks	Whether Filter reqd.	unit	Validation
1	Purpose	CH	Device for close wound drainage under negative pressure post opratively										YES
2	Size/ Colour Code	CH	6 FG LIGHT GREEN	8 FG BLUE	10 FG BLACK	12 FG WHITE	14 FG GREEN	16 FG ORANGE	18 FG RED		YES	FG	
3	Bellow Capacity	N	30	50							YES	ml	
4	Redon drain catheters having multi perforation for efficient drainage with Radio-opaque line and satin smooth eyes	CH	YES										
5	Connecting tube is Kink resistant and is provided with additional strength to withstand suction	CH	YES										
6	Easy to depress chamber to activate the suction of Bellow unit by one person	CH	YES										
7	Unit consist of Bellow unit, Needle,Connecting tube and Perforated Catheter	CH	YES										
8	Non Toxic/Pyrogen free/ Sterlized/Disposable/Individual Packed in peel off pouch	CH	YES										
	Good Manufacturing Practices(GMP) issued under Revised schedule-M of Drugs and Cosmetic Act 1940 as amended or WHO-Good Manufacturing Practices(GMP)	CH								Must Declare			
	Good Manufacturing Practices(GMP) Certificate No.	CH								Must Declare			
	Good Manufacturing Practices(GMP) Certificate date	CH								Must Declare			

9	Certification Available	CH	YES							Indicate the availability of Certification like ISO/Europeon CE/US-FDA	YES		
---	-------------------------	----	-----	--	--	--	--	--	--	--	-----	--	--

Approved By

(Sh.P.SINGHAL)DDG(Q.A.)(H.QRS)

Moderated By

(Sh.S.K.AGRAWAL)D(QA)

Moderated By

(Sh. A.K.M.KASHYAP)DD(QA)

Prepared By

(L.K.BHUGRA)AD(QA)