## FILE NO:DQA/DEL/MEC/GeM/TP/158

Sub Product Category -Laboratory mixing and stirring and shaking equipment and supplies UNSPSC CODE: 411038

Product category - Reciprocal shakers (Mechanical shaker machine with timer) UNSPSC CODE: 41103812 ITC-HS Code: 84798200.

(\*Demand aggregation to be done on UNSPSC code) S No. **Parameters** Type of Value 1 Remarks unit Field Purpose: Kahn racks, bottle or conical СН flasks of different sizes cab be shaken by yes adjusting the compartments provided on the shaking rack. Platform Length Ν Must cm Declare Platform Width Ν Must cm Declare Platform Width Ν Must cm Declare Material: mild steel sheet duly power СН coated and top surfaqce covered with non yes Heavy duty electric motor for continuous СН Must ves rating. Declare Shaking CH Reciprocating type Speed measurement. СН Digital OPM/speed Must meter. Declare Proximity switch. СН yes 10 СН Automatic digital Shaking time control. ON/OFF timer . 11 Speed range. Variable from 70 to 300 СН oscillations per minute. 12 Power Requirements. СН 230 V ± 10%, 50 Hz AC. 13 Power consumption Must СН watts Declare Availability of Test Reports of the product from Central Govt/NABL accredited Lab CH YES showing compliance to the specification Test report and approvals to be furnished СН YES to the buyer on demand. 16 Test Reports No. Must СН Declare 17 Test Reports Date Must CH Declare 18 Name of the Labs Must CH

Declare